

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PENNSAUKEN TWSP. P.D.
Serial No.: ARUM-0051
Calibration File No.: 03155 Calib. Date: 12/04/2023 Calib. No.: 00049
Certification File No.: 03156 Cert. Date: 12/04/2023 Cert. No.: 00043
Linearity File No.: 03157 Lin. Date: 12/04/2023 Lin. No.: 00042
Solution File No.: 03192 Soln. Date: 02/26/2024 Soln. No.: 00386
Sequential File No.: 03192 File Date: 02/26/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 04/04/2024
Solution Control Lot: 22110 Bottle No.: 1126

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:28S		
Control 1 EC	0.104%	21:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	21:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29S		
Control 2 EC	0.101%	21:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:31S		
Control 3 EC	0.101%	21:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	21:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARTINEZ

First Name: JUAN

MI: C

Signature: 

Badge No.: 27223

Date: 02/26/2024