

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02832	Cert. No.: 00032
Sequential File No.:	02832	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 12/30/2018
		Soln. No.: 00304
		File Date: 12/30/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0085

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:30S		
Control 1 EC	0.100%	08:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:31S		
Control 2 EC	0.099%	08:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		
Control 3 EC	0.098%	08:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: TIMOTHY

MI: 1

Signature: _____



Badge No.: 219

Date: 12/30/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02821	Cert. No.: 00032
Sequential File No.:	02821	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 12/06/2018
		Soln. No.: 00301
		File Date: 12/06/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0338
Control Solution %:	0.100%		Expires: 03/13/2020
Solution Control Lot:	18090		Bottle No.: 0023

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:14S		
Control 1 EC	0.099%	11:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16S		
Control 2 EC	0.098%	11:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17S		
Control 3 EC	0.099%	11:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CIACCIA

First Name: MICHAEL

MI: -

Signature: _____

Michael Ciaccia #172

Badge No.: 172

Date: 12/06/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02813	Cert. No.: 00032
Sequential File No.:	02813	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 11/16/2018
		Soln. No.: 00300
		File Date: 11/16/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0139

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:14S		
Control 1 EC	0.099%	10:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15S		
Control 2 EC	0.098%	10:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16S		
Control 3 EC	0.098%	10:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

Sgt M Henkel 149

Badge No.: 149

Date: 11/16/2018

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02812	Cert. No.: 00032
Sequential File No.:	02812	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 11/16/2018
		Soln. No.: 00299
		File Date: 11/16/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0704

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:01S		
Control 1 EC	-.---%	05:02S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	05:02S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PASQUALE

First Name: JODY

MI:

Signature: _____



Badge No.: 199

Date: 11/16/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02733	Calib. Date: 05/16/2018
Certification File No.:	02734	Calib. No.: 00035
Linearity File No.:	02735	Cert. Date: 05/16/2018
Solution File No.:	02798	Cert. No.: 00031
Sequential File No.:	02798	Lin. Date: 05/16/2018
		Lin. No.: 00030
		Soln. Date: 10/13/2018
		Soln. No.: 00297
		File Date: 10/13/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0715

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:32D		
Control 1 EC	0.099%	02:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:33D		
Control 2 EC	0.099%	02:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:34D		
Control 3 EC	0.099%	02:35D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PASQUALE

First Name: JODY

MI: P

Signature: _____



Badge No.: 199

Date: 10/13/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02733	Calib. Date: 05/16/2018
Certification File No.:	02734	Calib. No.: 00035
Linearity File No.:	02735	Cert. Date: 05/16/2018
Solution File No.:	02789	Cert. No.: 00031
Sequential File No.:	02789	Lin. Date: 05/16/2018
		Lin. No.: 00030
		Soln. Date: 09/17/2018
		Soln. No.: 00296
		File Date: 09/17/2018

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0339
Control Solution %:	0.100%			Expires:	03/13/2020
Solution Control Lot:	18090			Bottle No.:	0016

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:32D		
Control 1 EC	0.097%	07:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:33D		
Control 2 EC	0.097%	07:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:35D		
Control 3 EC	0.098%	07:35D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Badge No.: 216

Date: 09/17/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02733	Calib. Date: 05/16/2018
Certification File No.:	02734	Calib. No.: 00035
Linearity File No.:	02735	Cert. Date: 05/16/2018
Solution File No.:	02779	Cert. No.: 00031
Sequential File No.:	02779	Lin. Date: 05/16/2018
		Lin. No.: 00030
		Soln. Date: 09/02/2018
		Soln. No.: 00295
		File Date: 09/02/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0201

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:05D		
Control 1 EC	0.098%	10:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06D		
Control 2 EC	0.098%	10:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07D		
Control 3 EC	0.097%	10:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERN

First Name: RAYMOND

MI: T

Signature: _____

R. Kern 181

Badge No.: 181

Date: 09/02/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02733	Calib. Date: 05/16/2018
Certification File No.:	02734	Calib. No.: 00035
Linearity File No.:	02735	Cert. Date: 05/16/2018
Solution File No.:	02772	Cert. No.: 00031
Sequential File No.:	02772	Lin. Date: 05/16/2018
		Lin. No.: 00030
		Soln. Date: 08/18/2018
		Soln. No.: 00294
		File Date: 08/18/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0339
Control Solution %:	0.100%		Expires: 10/17/2019
Solution Control Lot:	17360		Bottle No.: 0883

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:20D		
Control 1 EC	0.099%	06:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		
Control 2 EC	0.099%	06:22D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		
Control 3 EC	0.099%	06:23D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 08/18/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02733	Calib. Date: 05/16/2018
Certification File No.:	02734	Calib. No.: 00035
Linearity File No.:	02735	Cert. Date: 05/16/2018
Solution File No.:	02763	Cert. No.: 00031
Sequential File No.:	02763	Lin. Date: 05/16/2018
		Lin. No.: 00030
		Soln. Date: 07/24/2018
		Soln. No.: 00293
		File Date: 07/24/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0339
Control Solution %:	0.100%		Expires: 10/17/2019
Solution Control Lot:	17360		Bottle No.: 0771

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:27D		
Control 1 EC	0.098%	03:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:29D		
Control 2 EC	0.098%	03:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:30D		
Control 3 EC	0.097%	03:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 191

Date: 07/24/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02733	Calib. Date: 05/16/2018
Certification File No.:	02734	Calib. No.: 00035
Linearity File No.:	02735	Cert. Date: 05/16/2018
Solution File No.:	02741	Cert. No.: 00031
Sequential File No.:	02741	Lin. Date: 05/16/2018
		Lin. No.: 00030
		Soln. Date: 06/17/2018
		Soln. No.: 00291
		File Date: 06/17/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0339
Control Solution %:	0.100%		Expires: 10/17/2019
Solution Control Lot:	17360		Bottle No.: 0824

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:20D		
Control 1 EC	0.100%	06:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21D		
Control 2 EC	0.099%	06:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		
Control 3 EC	0.099%	06:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HARTNETT

First Name: JAMES

MI: /

Signature: _____



Badge No.: 160

Date: 06/17/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02684	Calib. Date: 11/27/2017
Certification File No.:	02685	Calib. No.: 00034
Linearity File No.:	02686	Cert. Date: 11/27/2017
Solution File No.:	02732	Cert. No.: 00030
Sequential File No.:	02732	Lin. Date: 11/27/2017
		Lin. No.: 00029
		Soln. Date: 05/12/2018
		Soln. No.: 00289
		File Date: 05/12/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0772

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:25D		
Control 1 EC	0.098%	07:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26D		
Control 2 EC	0.097%	07:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28D		
Control 3 EC	0.098%	07:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 05/12/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02684	Calib. Date: 11/27/2017
Certification File No.:	02685	Calib. No.: 00034
Linearity File No.:	02686	Cert. Date: 11/27/2017
Solution File No.:	02724	Cert. No.: 00030
Sequential File No.:	02724	Lin. Date: 11/27/2017
		Lin. No.: 00029
		Soln. Date: 04/14/2018
		Soln. No.: 00288
		File Date: 04/14/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0805

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:17D		
Control 1 EC	0.099%	05:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:18D		
Control 2 EC	0.099%	05:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:20D		
Control 3 EC	0.099%	05:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 04/14/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02684	Calib. Date:	11/27/2017
Certification File No.:	02685	Cert. No.:	00034
Linearity File No.:	02686	Cert. Date:	11/27/2017
Solution File No.:	02717	Lin. No.:	00029
Sequential File No.:	02717	Lin. Date:	11/27/2017
		Soln. Date:	03/16/2018
		Soln. No.:	00287
		File Date:	03/16/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	17360	Expires:	10/17/2019
		Bottle No.:	0884

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:58S		
Control 1 EC	0.100%	07:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00S		
Control 2 EC	0.098%	08:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01S		
Control 3 EC	0.098%	08:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERN

First Name: RAYMOND

MI: T

Signature: _____

Ray Kern #181

Badge No.: 181

Date: 03/16/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02684	Calib. Date: 11/27/2017
Certification File No.:	02685	Calib. No.: 00034
Linearity File No.:	02686	Cert. Date: 11/27/2017
Solution File No.:	02710	Cert. No.: 00030
Sequential File No.:	02710	Lin. Date: 11/27/2017
		Lin. No.: 00029
		Soln. Date: 02/20/2018
		Soln. No.: 00286
		File Date: 02/20/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0878

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:24S		
Control 1 EC	0.099%	22:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:25S		
Control 2 EC	0.097%	22:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27S		
Control 3 EC	0.097%	22:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

Sgt M Henkel 149

Badge No.: 149

Date: 02/20/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02684	Calib. Date: 11/27/2017
Certification File No.:	02685	Calib. No.: 00034
Linearity File No.:	02686	Cert. Date: 11/27/2017
Solution File No.:	02705	Cert. No.: 00030
Sequential File No.:	02705	Lin. Date: 11/27/2017
		Lin. No.: 00029
		Soln. Date: 02/15/2018
		Soln. No.: 00285
		File Date: 02/15/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0770

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:59S		
Control 1 EC	0.100%	01:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:01S		
Control 2 EC	0.097%	01:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:02S		
Control 3 EC	0.098%	01:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 191

Date: 02/15/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02684	Calib. Date: 11/27/2017
Certification File No.:	02685	Calib. No.: 00034
Linearity File No.:	02686	Cert. Date: 11/27/2017
Solution File No.:	02699	Cert. No.: 00030
Sequential File No.:	02699	Lin. Date: 11/27/2017
		Lin. No.: 00029
		Soln. Date: 01/26/2018
		Soln. No.: 00284
		File Date: 01/26/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0955

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:26S		
Control 1 EC	0.099%	02:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:28S		
Control 2 EC	0.097%	02:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:29S		
Control 3 EC	0.097%	02:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 01/26/2018