

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02469 Calib. Date: 10/24/2018 Calib. No.: 00034
Certification File No.: 02470 Cert. Date: 10/24/2018 Cert. No.: 00030
Linearity File No.: 02471 Lin. Date: 10/24/2018 Lin. No.: 00030
Solution File No.: 02513 Soln. Date: 12/30/2018 Soln. No.: 00263
Sequential File No.: 02513 File Date: 12/30/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 03/13/2020
Solution Control Lot: 18090 Bottle No.: 0729

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:10S		
Control 1 EC	0.100%	08:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:12S		
Control 2 EC	0.099%	08:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13S		
Control 3 EC	0.100%	08:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CIACCIA

First Name: MICHAEL

MI: -

Signature:  #172

Badge No.: 172

Date: 12/30/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02469	Calib. Date: 10/24/2018
Certification File No.:	02470	Calib. No.: 00034
Linearity File No.:	02471	Cert. Date: 10/24/2018
Solution File No.:	02501	Cert. No.: 00030
Sequential File No.:	02501	Lin. Date: 10/24/2018
		Lin. No.: 00030
		Soln. Date: 12/13/2018
		Soln. No.: 00262
		File Date: 12/13/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0719

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:19S		
Control 1 EC	0.100%	16:19S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:20S		
Control 2 EC	0.098%	16:21S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:21S		
Control 3 EC	0.099%	16:22S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

Sgt M Henkel 149

Badge No.: 149

Date: 12/13/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02469	Calib. Date: 10/24/2018
Certification File No.:	02470	Calib. No.: 00034
Linearity File No.:	02471	Cert. Date: 10/24/2018
Solution File No.:	02486	Cert. No.: 00030
Sequential File No.:	02486	Lin. Date: 10/24/2018
		Lin. No.: 00030
		Soln. Date: 11/14/2018
		Soln. No.: 00261
		File Date: 11/14/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0341
Control Solution %:	0.100%		Expires: 03/13/2020
Solution Control Lot:	18090		Bottle No.: 0702

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:01S		
Control 1 EC	0.100%	03:01S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:02S		
Control 2 EC	0.099%	03:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:03S		
Control 3 EC	0.099%	03:04S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HURLEY

First Name: R

MI: J

Signature: P.H. R. #205

Badge No.: 205

Date: 11/14/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02422	Calib. Date: 05/16/2018
Certification File No.:	02423	Calib. No.: 00033
Linearity File No.:	02424	Cert. Date: 05/16/2018
Solution File No.:	02468	Cert. No.: 00029
Sequential File No.:	02468	Lin. Date: 05/16/2018
		Lin. No.: 00029
		Soln. Date: 10/23/2018
		Soln. No.: 00259
		File Date: 10/23/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %:	0.100%		Expires: 03/13/2020
Solution Control Lot:	18090		Bottle No.: 0760

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:32D		
Control 1 EC	0.098%	01:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:34D		
Control 2 EC	0.098%	01:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:35D		
Control 3 EC	0.098%	01:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Badge No.: 216

Date: 10/23/2018

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C		Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.		
Calibration File No.: 02422	Calib. Date: 05/16/2018	Calib. No.: 00033
Certification File No.: 02423	Cert. Date: 05/16/2018	Cert. No.: 00029
Linearity File No.: 02424	Lin. Date: 05/16/2018	Lin. No.: 00029
Solution File No.: 02459	Soln. Date: 09/17/2018	Soln. No.: 00257
Sequential File No.: 02459	File Date: 09/17/2018	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %: 0.100%		Expires: 03/13/2020
Solution Control Lot: 18090		Bottle No.: 0373

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:31D		
Control 1 EC	0.097%	07:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	07:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:32D		
Control 2 EC	0.096%	07:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	07:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:34D		
Control 3 EC	0.096%	07:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	07:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Badge No.: 216

Date: 09/17/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02422	Calib. Date: 05/16/2018
Certification File No.:	02423	Calib. No.: 00033
Linearity File No.:	02424	Cert. Date: 05/16/2018
Solution File No.:	02454	Cert. No.: 00029
Sequential File No.:	02454	Lin. Date: 05/16/2018
		Lin. No.: 00029
		Soln. Date: 09/02/2018
		Soln. No.: 00256
		File Date: 09/02/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0755

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:01D		
Control 1 EC	0.097%	10:01D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 2 EC	0.097%	10:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		
Control 3 EC	0.097%	10:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REESE

First Name: STEVEN

MI: R

Signature: _____



Badge No.: 137

Date: 09/02/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02422	Calib. Date: 05/16/2018
Certification File No.:	02423	Calib. No.: 00033
Linearity File No.:	02424	Cert. Date: 05/16/2018
Solution File No.:	02450	Cert. No.: 00029
Sequential File No.:	02450	Lin. Date: 05/16/2018
		Lin. No.: 00029
		Soln. Date: 08/18/2018
		Soln. No.: 00255
		File Date: 08/18/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0613

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:23D		
Control 1 EC	0.098%	06:24D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:24D		
Control 2 EC	0.098%	06:25D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	06:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25D		
Control 3 EC	0.097%	06:26D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Joe Kuchmek #216

Badge No.: 216

Date: 08/18/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02422	Calib. Date: 05/16/2018 Calib. No.: 00033
Certification File No.:	02423	Cert. Date: 05/16/2018 Cert. No.: 00029
Linearity File No.:	02424	Lin. Date: 05/16/2018 Lin. No.: 00029
Solution File No.:	02447	Soln. Date: 07/24/2018 Soln. No.: 00254
Sequential File No.:	02447	File Date: 07/24/2018
Calibrating Unit:	WET	Model No.: CU-34 Serial No.: DDUN S3-0340
Control Solution %:	0.100%	Expires: 10/17/2019
Solution Control Lot:	17360	Bottle No.: 0828

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:28D		
Control 1 EC	0.098%	01:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:29D		
Control 2 EC	0.097%	01:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:30D		
Control 3 EC	0.097%	01:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	01:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: _____

Badge No.: 191

Date: 07/24/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02422	Calib. Date: 05/16/2018
Certification File No.:	02423	Calib. No.: 00033
Linearity File No.:	02424	Cert. Date: 05/16/2018
Solution File No.:	02443	Cert. No.: 00029
Sequential File No.:	02443	Lin. Date: 05/16/2018
		Lin. No.: 00029
		Soln. Date: 06/29/2018
		Soln. No.: 00253
		File Date: 06/29/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %:	0.100%		Expires: 10/17/2019
Solution Control Lot:	17360		Bottle No.: 0834

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:30D		
Control 1 EC	0.097%	11:31D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		
Control 2 EC	0.097%	11:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33D		
Control 3 EC	0.097%	11:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Pr Joe # 216

Badge No.: 216

Date: 06/29/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02422	Calib. Date: 05/16/2018
Certification File No.:	02423	Calib. No.: 00033
Linearity File No.:	02424	Cert. Date: 05/16/2018
Solution File No.:	02438	Cert. No.: 00029
Sequential File No.:	02438	Lin. Date: 05/16/2018
		Lin. No.: 00029
		Soln. Date: 06/17/2018
		Soln. No.: 00252
		File Date: 06/17/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0822

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:21D		
Control 1 EC	0.097%	06:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		
Control 2 EC	0.096%	06:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	06:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		
Control 3 EC	0.096%	06:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PASQUALE

First Name: PTL

MI:

Signature: _____



Badge No.: 199

Date: 06/17/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02422	Calib. Date: 05/16/2018 Calib. No.: 00033
Certification File No.:	02423	Cert. Date: 05/16/2018 Cert. No.: 00029
Linearity File No.:	02424	Lin. Date: 05/16/2018 Lin. No.: 00029
Solution File No.:	02433	Soln. Date: 06/08/2018 Soln. No.: 00251
Sequential File No.:	02433	File Date: 06/08/2018
Calibrating Unit:	WET	Model No.: CU-34 Serial No.: DDUN S3-0340
Control Solution %:	0.100%	Expires: 10/17/2019
Solution Control Lot:	17360	Bottle No.: 0991

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:26D		
Control 1 EC	0.097%	22:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27D		
Control 2 EC	0.097%	22:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:29D		
Control 3 EC	0.097%	22:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 06/08/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02340	Calib. Date: 11/27/2017
Certification File No.:	02341	Calib. No.: 00032
Linearity File No.:	02342	Cert. Date: 11/27/2017
Solution File No.:	02418	Cert. No.: 00028
Sequential File No.:	02418	Lin. Date: 11/27/2017
		Lin. No.: 00028
		Soln. Date: 05/12/2018
		Soln. No.: 00249
		File Date: 05/12/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0341
Control Solution %:	0.100%		Expires: 10/17/2019
Solution Control Lot:	17360		Bottle No.: 0662

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:17D		
Control 1 EC	0.099%	07:17D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:18D		
Control 2 EC	0.098%	07:18D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:19D		
Control 3 EC	0.099%	07:20D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Pr Joe #216

Badge No.: 216

Date: 05/12/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02340	Calib. Date: 11/27/2017
Certification File No.:	02341	Calib. No.: 00032
Linearity File No.:	02342	Cert. Date: 11/27/2017
Solution File No.:	02407	Cert. No.: 00028
Sequential File No.:	02407	Lin. Date: 11/27/2017
		Lin. No.: 00028
		Soln. Date: 04/15/2018
		Soln. No.: 00248
		File Date: 04/15/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0341
Control Solution %:	0.100%		Expires: 10/17/2019
Solution Control Lot:	17360		Bottle No.: 0974

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:44D		
Control 1 EC	0.099%	06:45D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:45D		
Control 2 EC	0.098%	06:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:47D		
Control 3 EC	0.099%	06:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 04/15/2018

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02340	Calib. Date: 11/27/2017
Certification File No.:	02341	Calib. No.: 00032
Linearity File No.:	02342	Cert. Date: 11/27/2017
Solution File No.:	02392	Cert. No.: 00028
Sequential File No.:	02392	Lin. Date: 11/27/2017
		Lin. No.: 00028
		Soln. Date: 03/24/2018
		Soln. No.: 00247
		File Date: 03/24/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 1260

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:23S		
Control 1 EC	0.099%	04:23S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:24S		
Control 2 EC	0.098%	04:25S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:25S		
Control 3 EC	0.099%	04:26S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 191

Date: 03/24/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02340 Calib. Date: 11/27/2017 Calib. No.: 00032
Certification File No.: 02341 Cert. Date: 11/27/2017 Cert. No.: 00028
Linearity File No.: 02342 Lin. Date: 11/27/2017 Lin. No.: 00028
Solution File No.: 02382 Soln. Date: 03/16/2018 Soln. No.: 00246
Sequential File No.: 02382 File Date: 03/16/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 10/17/2019
Solution Control Lot: 17360 Bottle No.: 0891

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:36S		
Control 1 EC	0.097%	07:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	07:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		
Control 2 EC	0.097%	07:38S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39S		
Control 3 EC	0.097%	07:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERN

First Name: RAYMOND

MI: T

Signature: _____



Badge No.: 181

Date: 03/16/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02340	Calib. Date: 11/27/2017
Certification File No.:	02341	Calib. No.: 00032
Linearity File No.:	02342	Cert. Date: 11/27/2017
Solution File No.:	02374	Cert. No.: 00028
Sequential File No.:	02374	Lin. Date: 11/27/2017
		Lin. No.: 00028
		Soln. Date: 02/28/2018
		Soln. No.: 00245
		File Date: 02/28/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0903

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:25S		
Control 1 EC	0.100%	00:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:26S		
Control 2 EC	0.099%	00:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:28S		
Control 3 EC	0.098%	00:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 02/28/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02340	Calib. Date: 11/27/2017
Certification File No.:	02341	Calib. No.: 00032
Linearity File No.:	02342	Cert. Date: 11/27/2017
Solution File No.:	02359	Cert. No.: 00028
Sequential File No.:	02359	Lin. Date: 11/27/2017
		Lin. No.: 00028
		Soln. Date: 01/26/2018
		Soln. No.: 00244
		File Date: 01/26/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0341
Control Solution %:	0.100%		Expires: 10/17/2019
Solution Control Lot:	17360		Bottle No.: 0827

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:23S		
Control 1 EC	0.099%	02:24S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:25S		
Control 2 EC	0.098%	02:25S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:26S		
Control 3 EC	0.099%	02:27S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Joseph #216

Badge No.: 216

Date: 01/26/2018