

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02750	Calib. Date: 06/25/2020
Certification File No.:	02751	Calib. No.: 00038
Linearity File No.:	02752	Cert. Date: 06/25/2020
Solution File No.:	02794	Cert. No.: 00034
Sequential File No.:	02794	Lin. Date: 06/25/2020
		Lin. No.: 00034
		Soln. Date: 11/27/2020
		Soln. No.: 00299
		File Date: 11/27/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1301

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:57S		
Control 1 EC	0.101%	21:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:58S		
Control 2 EC	0.100%	21:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:59S		
Control 3 EC	0.101%	22:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MIRELES

First Name: T

MI: -

Signature: \_\_\_\_\_

Badge No.: 219

Date: 11/27/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02750	Calib. Date: 06/25/2020
Certification File No.:	02751	Calib. No.: 00038
Linearity File No.:	02752	Cert. Date: 06/25/2020
Solution File No.:	02789	Cert. No.: 00034
Sequential File No.:	02789	Lin. Date: 06/25/2020
		Lin. No.: 00034
		Soln. Date: 10/31/2020
		Soln. No.: 00298
		File Date: 10/31/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1275

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:12S		
Control 1 EC	0.101%	05:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:13S		
Control 2 EC	0.100%	05:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	05:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:15S		
Control 3 EC	0.100%	05:15S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MIRELES

First Name: T

MI: -

Signature: \_\_\_\_\_



Badge No.: 219

Date: 10/31/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02750	Calib. Date: 06/25/2020
Certification File No.:	02751	Calib. No.: 00038
Linearity File No.:	02752	Cert. Date: 06/25/2020
Solution File No.:	02777	Cert. No.: 00034
Sequential File No.:	02777	Lin. Date: 06/25/2020
		Lin. No.: 00034
		Soln. Date: 10/03/2020
		Soln. No.: 00297
		File Date: 10/03/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1312

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:55D		
Control 1 EC	0.101%	02:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:56D		
Control 2 EC	0.100%	02:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:58D		
Control 3 EC	0.099%	02:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MIRELES

First Name: T

MI: -

Signature: 

Badge No.: 219

Date: 10/03/2020

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02750 Calib. Date: 06/25/2020 Calib. No.: 00038  
Certification File No.: 02751 Cert. Date: 06/25/2020 Cert. No.: 00034  
Linearity File No.: 02752 Lin. Date: 06/25/2020 Lin. No.: 00034  
Solution File No.: 02769 Soln. Date: 08/29/2020 Soln. No.: 00296  
Sequential File No.: 02769 File Date: 08/29/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 09/09/2021  
Solution Control Lot: 19240 Bottle No.: 1355

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:05D		
Control 1 EC	0.101%	04:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:06D		
Control 2 EC	0.100%	04:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:07D		
Control 3 EC	0.100%	04:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PASQUALE

First Name: JODY

MI:

Signature: \_\_\_\_\_



Badge No.: 199

Date: 08/29/2020

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02750 Calib. Date: 06/25/2020 Calib. No.: 00038  
Certification File No.: 02751 Cert. Date: 06/25/2020 Cert. No.: 00034  
Linearity File No.: 02752 Lin. Date: 06/25/2020 Lin. No.: 00034  
Solution File No.: 02761 Soln. Date: 08/13/2020 Soln. No.: 00295  
Sequential File No.: 02761 File Date: 08/13/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 09/09/2021  
Solution Control Lot: 19240 Bottle No.: 1303

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:18D		
Control 1 EC	0.099%	19:18D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:19D		
Control 2 EC	0.098%	19:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:20D		
Control 3 EC	0.099%	19:21D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: HURLEY

First Name: R

MI:

Signature: PH. R. H. # 205

Badge No.: 205

Date: 08/13/2020

**Calibrating Unit**  
**New Standard Solution Report**

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02750 Calib. Date: 06/25/2020 Calib. No.: 00038  
Certification File No.: 02751 Cert. Date: 06/25/2020 Cert. No.: 00034  
Linearity File No.: 02752 Lin. Date: 06/25/2020 Lin. No.: 00034  
Solution File No.: 02755 Soln. Date: 07/26/2020 Soln. No.: 00294  
Sequential File No.: 02755 File Date: 07/26/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341  
Control Solution %: 0.100% Expires: 09/09/2021  
Solution Control Lot: 19240 Bottle No.: 1366

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:41D		
Control 1 EC	0.102%	03:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:42D		
Control 2 EC	0.101%	03:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:43D		
Control 3 EC	0.101%	03:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 191

Date: 07/26/2020

**Calibrating Unit**  
**New Standard Solution Report**

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02714 Calib. Date: 01/16/2020 Calib. No.: 00037  
Certification File No.: 02715 Cert. Date: 01/16/2020 Cert. No.: 00033  
Linearity File No.: 02716 Lin. Date: 01/16/2020 Lin. No.: 00033  
Solution File No.: 02747 Soln. Date: 06/11/2020 Soln. No.: 00292  
Sequential File No.: 02747 File Date: 06/11/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340  
Control Solution %: 0.100% Expires: 09/09/2021  
Solution Control Lot: 19240 Bottle No.: 1362

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:24D		
Control 1 EC	0.100%	05:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:26D		
Control 2 EC	0.099%	05:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	05:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:27D		
Control 3 EC	0.099%	05:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HURLEY

First Name: R

MI:

Signature: \_\_\_\_\_

*PH. R. H. #205*

Badge No.: 205

Date: 06/11/2020

**Calibrating Unit**  
**New Standard Solution Report**

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02714 Calib. Date: 01/16/2020 Calib. No.: 00037  
Certification File No.: 02715 Cert. Date: 01/16/2020 Cert. No.: 00033  
Linearity File No.: 02716 Lin. Date: 01/16/2020 Lin. No.: 00033  
Solution File No.: 02740 Soln. Date: 05/11/2020 Soln. No.: 00291  
Sequential File No.: 02740 File Date: 05/11/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340  
Control Solution %: 0.100% Expires: 09/09/2021  
Solution Control Lot: 19240 Bottle No.: 1380

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:01D		
Control 1 EC	0.100%	10:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 2 EC	0.098%	10:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		
Control 3 EC	0.098%	10:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: PETER

MI: E

Signature: \_\_\_\_\_



Badge No.: 187

Date: 05/11/2020



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02714	Calib. Date: 01/16/2020
Certification File No.:	02715	Calib. No.: 00037
Linearity File No.:	02716	Cert. Date: 01/16/2020
Solution File No.:	02735	Cert. No.: 00033
Sequential File No.:	02735	Lin. Date: 01/16/2020
		Lin. No.: 00033
		Soln. Date: 04/11/2020
		Soln. No.: 00290
		File Date: 04/11/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1372

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:57D		
Control 1 EC	0.099%	01:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:58D		
Control 2 EC	0.099%	01:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:00D		
Control 3 EC	0.099%	02:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MIRELES

First Name: T

MI: -

Signature: \_\_\_\_\_

*P.H. Mireles #219*

Badge No.: 219

Date: 04/11/2020

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02714	Calib. Date: 01/16/2020
Certification File No.:	02715	Calib. No.: 00037
Linearity File No.:	02716	Cert. Date: 01/16/2020
Solution File No.:	02731	Cert. No.: 00033
Sequential File No.:	02731	Lin. Date: 01/16/2020
		Lin. No.: 00033
		Soln. Date: 03/13/2020
		Soln. No.: 00289
		File Date: 03/13/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %:	0.100%		Expires: 06/10/2021
Solution Control Lot:	19150		Bottle No.: 0311

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:23S		●
Control 1 EC	0.099%	23:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:25S		
Control 2 EC	0.098%	23:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:26S		
Control 3 EC	0.098%	23:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MIRELES

First Name: T

MI: -

Signature: \_\_\_\_\_

*M. Mireles*  
#219

Badge No.: 219

Date: 03/13/2020

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02714	Calib. Date: 01/16/2020
Certification File No.:	02715	Calib. No.: 00037
Linearity File No.:	02716	Cert. Date: 01/16/2020
Solution File No.:	02722	Cert. No.: 00033
Sequential File No.:	02722	Lin. Date: 01/16/2020
		Lin. No.: 00033
		Soln. Date: 02/13/2020
		Soln. No.: 00288
		File Date: 02/13/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0261

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:19S		
Control 1 EC	0.099%	23:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:21S		
Control 2 EC	0.098%	23:21S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:22S		
Control 3 EC	0.099%	23:23S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: \_\_\_\_\_



Badge No.: 149

Date: 02/13/2020

**Calibrating Unit**  
**New Standard Solution Report**

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02714 Calib. Date: 01/16/2020 Calib. No.: 00037  
Certification File No.: 02715 Cert. Date: 01/16/2020 Cert. No.: 00033  
Linearity File No.: 02716 Lin. Date: 01/16/2020 Lin. No.: 00033  
Solution File No.: 02722 Soln. Date: 02/13/2020 Soln. No.: 00288  
Sequential File No.: 02722 File Date: 02/13/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340  
Control Solution %: 0.100% Expires: 06/10/2021  
Solution Control Lot: 19150 Bottle No.: 0261

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:19S		
Control 1 EC	0.099%	23:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:21S		
Control 2 EC	0.098%	23:21S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:22S		
Control 3 EC	0.099%	23:23S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: LTJ Henkel 149

Badge No.: 149

Date: 02/13/2020



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02650	Calib. Date: 08/26/2019
Certification File No.:	02651	Calib. No.: 00036
Linearity File No.:	02652	Cert. Date: 08/26/2019
Solution File No.:	02712	Cert. No.: 00032
Sequential File No.:	02712	Lin. Date: 08/26/2019
		Lin. No.: 00032
		Soln. Date: 01/11/2020
		Soln. No.: 00286
		File Date: 01/11/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1374

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:40S		
Control 1 EC	0.100%	20:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:41S		
Control 2 EC	0.100%	20:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:43S		
Control 3 EC	0.100%	20:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MIRELES

First Name: T

MI: -

Signature: \_\_\_\_\_

*PH. A. Mireles #219*

Badge No.: 219

Date: 01/11/2020